


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041822\
 Data File : VN072027.D
 Acq On : 18 Apr 2022 10:31
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu Mar 24 12:03:44 2022



AutoFind: Averaged scan 1819 to 1836; Bkg corrected with scan 1818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.2	3428	PASS
75	95	30	60	46.6	10545	PASS
95	95	100	100	100.0	22607	PASS
96	95	5	9	7.1	1596	PASS
173	174	0.00	2	1.1	223	PASS
174	95	50	100	88.3	19951	PASS
175	174	5	9	7.4	1485	PASS
176	174	95	101	97.0	19348	PASS
177	176	5	9	6.4	1242	PASS