

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042020\
 Data File : VN061102.D
 Acq On : 20 Apr 2020 15:29
 Operator : JC/MD
 Sample : L2338-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-HN-MW24I-20200415

Quant Time: Apr 21 05:51:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	147197	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	253589	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	233196	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	100541	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101797	54.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.10%	
35) Dibromofluoromethane	7.56	113	76395	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
50) Toluene-d8	10.08	98	310420	55.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.94%	
62) 4-Bromofluorobenzene	12.40	95	112972	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
Target Compounds						
7) Trichlorofluoromethane	3.01	101	1731	0.641	ug/l	96
19) Methyl tert-butyl Ether	5.01	73	5623	1.001	ug/l	91
44) Trichloroethene	8.82	130	11929	6.115	ug/l	99
64) Tetrachloroethene	10.62	164	3939	2.207	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042020\
Data File : VN061102.D
Acq On : 20 Apr 2020 15:29
Operator : JC/MD
Sample : L2338-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-HN-MW24I-20200415

Quant Time: Apr 21 05:51:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

