

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042020\
 Data File : VN061104.D
 Acq On : 20 Apr 2020 16:14
 Operator : JC/MD
 Sample : L2330-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-SO-2020-04-16

Quant Time: Apr 21 05:56:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	145227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	252132	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	237351	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	102371	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101149	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	7.56	113	75772	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
50) Toluene-d8	10.08	98	311618	56.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.02%	
62) 4-Bromofluorobenzene	12.39	95	113576	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
Target Compounds						
16) Acetone	3.79	43	8651	10.578	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042020\
Data File : VN061104.D
Acq On : 20 Apr 2020 16:14
Operator : JC/MD
Sample : L2330-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-SO-2020-04-16

Quant Time: Apr 21 05:56:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

