

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042020\
 Data File : VN061108.D
 Acq On : 20 Apr 2020 17:45
 Operator : JC/MD
 Sample : L2330-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP-2020-04-16

Quant Time: Apr 21 06:07:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	144986	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	250277	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	237298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	104442	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	99996	53.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.82%	
35) Dibromofluoromethane	7.56	113	75125	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
50) Toluene-d8	10.08	98	311484	56.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.80%	
62) 4-Bromofluorobenzene	12.40	95	112066	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
Target Compounds						
95) Naphthalene	15.13	128	14852	2.589	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042020\
Data File : VN061108.D
Acq On : 20 Apr 2020 17:45
Operator : JC/MD
Sample : L2330-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP-2020-04-16

Quant Time: Apr 21 06:07:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

