

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042020\
 Data File : VN061109.D
 Acq On : 20 Apr 2020 18:08
 Operator : JC/MD
 Sample : L2330-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PA-TWP-1

Quant Time: Apr 21 06:09:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	145394	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	252078	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	237687	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	103565	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	101137	54.37	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	108.74%	
35) Dibromofluoromethane	7.56	113	76697	52.10	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.20%	
50) Toluene-d8	10.08	98	314406	56.52	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	113.04%	
62) 4-Bromofluorobenzene	12.40	95	113926	52.25	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.50%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042020\
Data File : VN061109.D
Acq On : 20 Apr 2020 18:08
Operator : JC/MD
Sample : L2330-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PA-TWP-1

Quant Time: Apr 21 06:09:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

