

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042020\
 Data File : VN061112.D
 Acq On : 20 Apr 2020 19:15
 Operator : JC/MD
 Sample : L2330-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WB-TWP-3

Quant Time: Apr 21 06:31:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	143046	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	244204	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	230255	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101653	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	98803	53.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.98%	
35) Dibromofluoromethane	7.56	113	75139	52.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.38%	
50) Toluene-d8	10.08	98	302025	56.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.08%	
62) 4-Bromofluorobenzene	12.40	95	108570	51.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042020\
Data File : VN061112.D
Acq On : 20 Apr 2020 19:15
Operator : JC/MD
Sample : L2330-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WB-TWP-3

Quant Time: Apr 21 06:31:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

