

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042020\
 Data File : VN061122.D
 Acq On : 21 Apr 2020 00:10
 Operator : JC/MD
 Sample : PB128313TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128313TB

Quant Time: Apr 21 06:56:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	147516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	251842	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	239653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	106324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	100977	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
35) Dibromofluoromethane	7.56	113	75456	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
50) Toluene-d8	10.08	98	314755	56.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.28%	
62) 4-Bromofluorobenzene	12.40	95	116374	53.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.86%	
Target Compounds						
16) Acetone	3.79	43	5870	7.066	ug/l	98
43) Isopropyl Acetate	8.14	43	58030	12.644	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042020\
Data File : VN061122.D
Acq On : 21 Apr 2020 00:10
Operator : JC/MD
Sample : PB128313TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128313TB

Quant Time: Apr 21 06:56:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

