


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042020\
 Data File : VN061117.D
 Acq On : 20 Apr 2020 21:54
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Title : SW846 8260
 Last Update : Fri Apr 17 03:49:39 2020



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	5839	PASS
75	95	30	60	52.9	15605	PASS
95	95	100	100	100.0	29525	PASS
96	95	5	9	6.9	2048	PASS
173	174	0.00	2	0.4	84	PASS
174	95	50	100	76.3	22514	PASS
175	174	5	9	8.2	1848	PASS
176	174	95	101	98.1	22088	PASS
177	176	5	9	7.0	1550	PASS