

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
 Data File : VN082182.D
 Acq On : 17 May 2024 16:34
 Operator : JC\MD
 Sample : PB160967TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160967TB

Quant Time: May 17 23:12:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134095	51.248	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.500%
35) Dibromofluoromethane	8.171	113	92631	51.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.840%
50) Toluene-d8	10.571	98	369501	54.018	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.040%
62) 4-Bromofluorobenzene	12.853	95	122641	50.232	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.460%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	16979	15.818	ug/l #	88
20) Methylene Chloride	5.277	84	8474	3.913	ug/l #	83
37) Ethyl Acetate	7.565	43	4510	1.347	ug/l #	84
43) Isopropyl Acetate	8.689	43	205949	34.708	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
Data File : VN082182.D
Acq On : 17 May 2024 16:34
Operator : JC\MD
Sample : PB160967TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160967TB

Quant Time: May 17 23:12:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

