

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
 Data File : VN082183.D
 Acq On : 17 May 2024 16:58
 Operator : JC\MD
 Sample : PB160967ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160967ZHE#01

Quant Time: May 17 23:13:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	317533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	145604	51.867	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.740%		
35) Dibromofluoromethane	8.165	113	103657	53.067	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.140%		
50) Toluene-d8	10.571	98	406203	54.241	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.480%		
62) 4-Bromofluorobenzene	12.847	95	133747	50.037	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 100.080%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	16667	14.473	ug/l	94
20) Methylene Chloride	5.277	84	6931	2.983	ug/l #	86
43) Isopropyl Acetate	8.688	43	251049	38.644	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
Data File : VN082183.D
Acq On : 17 May 2024 16:58
Operator : JC\MD
Sample : PB160967ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160967ZHE#01

Quant Time: May 17 23:13:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

