

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
 Data File : VN082184.D
 Acq On : 17 May 2024 17:22
 Operator : JC\MD
 Sample : PB160967ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160967ZHE#02

Quant Time: May 17 23:13:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	327417	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	308205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	154497	52.621	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.240%		
35) Dibromofluoromethane	8.165	113	107149	53.199	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.400%		
50) Toluene-d8	10.571	98	427156	55.317	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 110.640%		
62) 4-Bromofluorobenzene	12.853	95	141647	51.393	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 102.780%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	15127	12.560	ug/l #	85
20) Methylene Chloride	5.277	84	6676	2.747	ug/l #	73
43) Isopropyl Acetate	8.688	43	241722	36.085	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
Data File : VN082184.D
Acq On : 17 May 2024 17:22
Operator : JC\MD
Sample : PB160967ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160967ZHE#02

Quant Time: May 17 23:13:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

