

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
 Data File : VN082185.D
 Acq On : 17 May 2024 17:46
 Operator : JC\MD
 Sample : PB160967ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160967ZHE#03

Quant Time: May 17 23:14:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	313139	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	302032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	145555	53.274	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.540%	
35) Dibromofluoromethane	8.171	113	101527	52.706	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.420%	
50) Toluene-d8	10.571	98	407294	55.150	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.300%	
62) 4-Bromofluorobenzene	12.853	95	134315	50.955	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 101.900%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	15161	13.527	ug/l	91
20) Methylene Chloride	5.283	84	6302	2.787	ug/l #	77
43) Isopropyl Acetate	8.688	43	267049	41.684	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
Data File : VN082185.D
Acq On : 17 May 2024 17:46
Operator : JC\MD
Sample : PB160967ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160967ZHE#03

Quant Time: May 17 23:14:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

