

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
 Data File : VN082192.D
 Acq On : 17 May 2024 20:34
 Operator : JC\MD
 Sample : PB160967ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160967ZHE#10

Quant Time: May 17 23:17:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	170386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301290	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	119245	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	139193	51.736	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.480%		
35) Dibromofluoromethane	8.171	113	97381	52.542	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.080%		
50) Toluene-d8	10.571	98	396010	55.731	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 111.460%		
62) 4-Bromofluorobenzene	12.853	95	130348	51.395	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 102.780%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	17633	15.977	ug/l	99
20) Methylene Chloride	5.277	84	8048	3.614	ug/l #	87
37) Ethyl Acetate	7.565	43	4133	1.188	ug/l #	74
43) Isopropyl Acetate	8.688	43	290656	47.153	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
Data File : VN082192.D
Acq On : 17 May 2024 20:34
Operator : JC\MD
Sample : PB160967ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160967ZHE#10

Quant Time: May 17 23:17:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

