

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051724\  
 Data File : VN082193.D  
 Acq On : 17 May 2024 20:58  
 Operator : JC\MD  
 Sample : P2480-04  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 WC-MH-JJ

Quant Time: May 17 23:17:53 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051524W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 16 05:21:06 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.224	168	174628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	313403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	142512	51.683	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.360%
35) Dibromofluoromethane	8.165	113	100079	51.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.820%
50) Toluene-d8	10.571	98	403126	54.539	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.080%
62) 4-Bromofluorobenzene	12.853	95	134728	51.068	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	102.140%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	17379	15.364	ug/l	90
20) Methylene Chloride	5.277	84	14593	6.394	ug/l #	80
43) Isopropyl Acetate	8.688	43	257980	40.235	ug/l #	70
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051724\  
 Data File : VN082193.D  
 Acq On : 17 May 2024 20:58  
 Operator : JC\MD  
 Sample : P2480-04  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 WC-MH-JJ

Quant Time: May 17 23:17:53 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051524W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 16 05:21:06 2024  
 Response via : Initial Calibration

