

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
 Data File : VN082197.D
 Acq On : 17 May 2024 22:34
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: May 18 00:13:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	297519	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	276315	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117449	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	137742	52.568	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.140%
35) Dibromofluoromethane	8.165	113	96264	52.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.200%
50) Toluene-d8	10.571	98	377113	53.744	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.480%
62) 4-Bromofluorobenzene	12.847	95	126231	50.402	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051724\
Data File : VN082197.D
Acq On : 17 May 2024 22:34
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: May 18 00:13:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

