

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061418\
 Data File : VN049220.D
 Acq On : 14 Jun 2018 11:44
 Operator : MD\SY
 Sample : VN0614WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0614WBL01

Manual Integrations
 APPROVED

MMDadoda
 6/15/2018 1:15:19 PM

Quant Time: Jun 14 17:09:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 14 05:15:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	533783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1035501	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1118270	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	340219	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	385902	44.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.54%	
35) Dibromofluoromethane	7.59	113	348750	40.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.38%	
50) Toluene-d8	10.09	98	1324236	39.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.44%	
62) 4-Bromofluorobenzene	12.40	95	448420	40.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.04%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061418\
Data File : VN049220.D
Acq On : 14 Jun 2018 11:44
Operator : MD\SY
Sample : VN0614WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0614WBL01

Manual Integrations
APPROVED

MMDadoda
6/15/2018 1:15:19 PM

Quant Time: Jun 14 17:09:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 14 05:15:23 2018
Response via : Initial Calibration

