

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061418\
 Data File : VN049236.D
 Acq On : 14 Jun 2018 18:51
 Operator : MD\SY
 Sample : J3435-04DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW11DL

Manual Integrations
 APPROVED

MMDadoda
 6/15/2018 1:15:30 PM

Quant Time: Jun 15 11:36:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 14 05:15:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	600860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1111998	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1170560	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	447113	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	406344	41.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.76%	
35) Dibromofluoromethane	7.59	113	357091	38.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.60%	
50) Toluene-d8	10.09	98	1404044	39.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.42%	
62) 4-Bromofluorobenzene	12.40	95	539505	45.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.80%	
Target Compounds						
					Qvalue	
31) Cyclohexane	7.66	56	294916	20.217	ug/l	98
39) Methylcyclohexane	9.08	83	116061	8.087	ug/l	94
67) Ethyl Benzene	11.51	91	1335849	25.643	ug/l	100
68) m/p-Xylenes	11.63	106	859512	45.118	ug/l	99
69) o-Xylene	11.95	106	220774	12.034	ug/l	99
73) Isopropylbenzene	12.25	105	127411	2.967	ug/l	100
78) n-propylbenzene	12.59	91	360447	7.555	ug/l	97
80) 1,3,5-Trimethylbenzene	12.73	105	376549	11.436	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	1768421	54.033	ug/l	100
95) Naphthalene	15.14	128	139581	13.431	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061418\
Data File : VN049236.D
Acq On : 14 Jun 2018 18:51
Operator : MD\SY
Sample : J3435-04DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW11DL

Manual Integrations
APPROVED

MMDadoda
6/15/2018 1:15:30 PM

Quant Time: Jun 15 11:36:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 14 05:15:23 2018
Response via : Initial Calibration

