

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061418\
 Data File : VN049239.D
 Acq On : 14 Jun 2018 20:08
 Operator : MD\SY
 Sample : J3447-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 W-20

Quant Time: Jun 25 17:27:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 14 05:15:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	622819	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1201066	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1304174	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	384672	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	464629	46.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.40%	
35) Dibromofluoromethane	7.59	113	402120	40.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.90%	
50) Toluene-d8	10.09	98	1567065	40.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.04%	
62) 4-Bromofluorobenzene	12.41	95	535530	41.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.44%	
Target Compounds						
24) 1,1-Dichloroethane	5.85	63	21501	1.24	ug/l	98
85) sec-Butylbenzene	13.17	105	29186	0.90	ug/l	96
95) Naphthalene	15.14	128	6779	5.87	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061418\
Data File : VN049239.D
Acq On : 14 Jun 2018 20:08
Operator : MD\SY
Sample : J3447-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
W-20

Quant Time: Jun 25 17:27:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 14 05:15:23 2018
Response via : Initial Calibration

