

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061418\
 Data File : VN049240.D
 Acq On : 14 Jun 2018 20:34
 Operator : MD\SY
 Sample : J3447-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 W-19

Quant Time: Jun 15 06:14:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 14 05:15:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	646986	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1251218	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1331796	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	455968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	469566	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	
35) Dibromofluoromethane	7.59	113	415013	40.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.14%	
50) Toluene-d8	10.09	98	1636984	40.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.26%	
62) 4-Bromofluorobenzene	12.41	95	621635	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
Target Compounds						
24) 1,1-Dichloroethane	5.85	63	27956	1.549	ug/l #	93
44) Trichloroethene	8.84	130	9060	0.766	ug/l	96
83) tert-Butylbenzene	12.99	119	13061	0.452	ug/l #	60
85) sec-Butylbenzene	13.17	105	191077	4.955	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061418\
Data File : VN049240.D
Acq On : 14 Jun 2018 20:34
Operator : MD\SY
Sample : J3447-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
W-19

Quant Time: Jun 15 06:14:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 14 05:15:23 2018
Response via : Initial Calibration

