

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061418\
 Data File : VN049242.D
 Acq On : 14 Jun 2018 21:25
 Operator : MD\SY
 Sample : J3447-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 W-22

Quant Time: Jun 15 06:53:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 14 05:15:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	619744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1185477	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1200566	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	345585	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	454095	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
35) Dibromofluoromethane	7.59	113	399131	40.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.36%	
50) Toluene-d8	10.09	98	1503862	39.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.80%	
62) 4-Bromofluorobenzene	12.41	95	480517	37.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.86%	
Target Compounds						
16) Acetone	3.82	43	10498	4.390	ug/l #	88
30) Chloroform	7.37	83	5292	0.319	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061418\
Data File : VN049242.D
Acq On : 14 Jun 2018 21:25
Operator : MD\SY
Sample : J3447-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
W-22

Quant Time: Jun 15 06:53:38 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061318W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 14 05:15:23 2018
Response via : Initial Calibration

