

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049272.D
 Acq On : 15 Jun 2018 21:00
 Operator : MD\SY
 Sample : PB110255ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110255ZHE#03

Quant Time: Jun 16 02:06:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1262333	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1991230	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1635437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	428058	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	823195	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
35) Dibromofluoromethane	7.59	113	740622	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
50) Toluene-d8	10.09	98	2725705	44.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.88%	
62) 4-Bromofluorobenzene	12.41	95	657766	32.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.86%	
Target Compounds						
16) Acetone	3.82	43	46974	11.56	ug/l	93
18) Methyl Acetate	4.33	43	84528	6.82	ug/l	94
20) Methylene Chloride	4.55	84	28401	1.67	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049272.D
Acq On : 15 Jun 2018 21:00
Operator : MD\SY
Sample : PB110255ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110255ZHE#03

Quant Time: Jun 16 02:06:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

