

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049279.D
 Acq On : 16 Jun 2018 00:00
 Operator : MD\SY
 Sample : PB110255ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110255ZHE#10

Quant Time: Jun 16 02:29:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1316746	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2138639	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1616549	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	420915	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	887619	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
35) Dibromofluoromethane	7.59	113	797452	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
50) Toluene-d8	10.09	98	2835677	43.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.08%	
62) 4-Bromofluorobenzene	12.41	95	636145	29.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.40%	
Target Compounds						
16) Acetone	3.82	43	45076	10.63	ug/l	99
18) Methyl Acetate	4.33	43	123884	9.58	ug/l	94
20) Methylene Chloride	4.55	84	32220	1.82	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049279.D
Acq On : 16 Jun 2018 00:00
Operator : MD\SY
Sample : PB110255ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110255ZHE#10

Quant Time: Jun 16 02:29:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

