

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049281.D
 Acq On : 16 Jun 2018 00:52
 Operator : MD\SY
 Sample : PB110255ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110255ZHE#12

Quant Time: Jun 16 03:22:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1192319	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1929528	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1588903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	433444	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	798400	50.49	ug/l	0.00
Spiked Amount						
			Recovery	=	100.98%	
35) Dibromofluoromethane	7.59	113	720736	46.17	ug/l	0.00
Spiked Amount						
			Recovery	=	92.34%	
50) Toluene-d8	10.09	98	2637469	44.37	ug/l	0.00
Spiked Amount						
			Recovery	=	88.74%	
62) 4-Bromofluorobenzene	12.41	95	648174	32.98	ug/l	0.00
Spiked Amount						
			Recovery	=	65.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	40205	10.47	ug/l	93
18) Methyl Acetate	4.33	43	114899	9.81	ug/l	95
20) Methylene Chloride	4.55	84	23670	1.48	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049281.D
Acq On : 16 Jun 2018 00:52
Operator : MD\SY
Sample : PB110255ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110255ZHE#12

Quant Time: Jun 16 03:22:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

