

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049282.D
 Acq On : 16 Jun 2018 1:17
 Operator : MD\SY
 Sample : PB110255ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110255ZHE#13

Quant Time: Jun 16 03:24:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1201062	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1934232	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1527426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	349328	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	795774	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
35) Dibromofluoromethane	7.59	113	717120	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
50) Toluene-d8	10.09	98	2656715	44.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.18%	
62) 4-Bromofluorobenzene	12.40	95	573764	29.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.24%	
Target Compounds						
16) Acetone	3.82	43	41026	10.61	ug/l	# 84
18) Methyl Acetate	4.33	43	117812	9.99	ug/l	96
20) Methylene Chloride	4.55	84	24452	1.51	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049282.D
Acq On : 16 Jun 2018 1:17
Operator : MD\SY
Sample : PB110255ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110255ZHE#13

Quant Time: Jun 16 03:24:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

