

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049314.D
 Acq On : 16 Jun 2018 17:20
 Operator : MD\SY
 Sample : J3564-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MANHATTANVILLE-TRIP-BLANK-1-

Quant Time: Jun 18 10:01:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1443768	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2278706	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1916666	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	547102	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	970081	50.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.32%	
35) Dibromofluoromethane	7.59	113	869581	47.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.34%	
50) Toluene-d8	10.09	98	3196276	45.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.06%	
62) 4-Bromofluorobenzene	12.40	95	804688	34.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	69.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049314.D
Acq On : 16 Jun 2018 17:20
Operator : MD\SY
Sample : J3564-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MANHATTANVILLE-TRIP-BLANK-1-

Quant Time: Jun 18 10:01:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

