

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049316.D
 Acq On : 16 Jun 2018 18:12
 Operator : MD\SY
 Sample : J3564-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-20

Manual Integrations
 APPROVED

MMDadoda
 6/18/2018 3:16:26 PM

Quant Time: Jun 18 10:02:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1510895	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2411649	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2065096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	803360	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	1006244	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
35) Dibromofluoromethane	7.59	113	904856	46.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.76%	
50) Toluene-d8	10.09	98	3373796	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
62) 4-Bromofluorobenzene	12.40	95	1004269	40.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.76%	
Target Compounds						
39) Methylcyclohexane	9.08	83	9651	0.33	ug/l	95
73) Isopropylbenzene	12.25	105	220435	3.13	ug/l	99
78) n-propylbenzene	12.59	91	287744	3.72	ug/l	99
83) tert-Butylbenzene	13.00	119	38111	0.78	ug/l	93
85) sec-Butylbenzene	13.17	105	98354m	1.50	ug/l	
89) n-Butylbenzene	13.61	91	21500m	0.50	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049316.D
Acq On : 16 Jun 2018 18:12
Operator : MD\SY
Sample : J3564-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-20

Manual Integrations
APPROVED

MMDadoda
6/18/2018 3:16:26 PM

Quant Time: Jun 18 10:02:46 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

