

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049326.D
 Acq On : 16 Jun 2018 22:29
 Operator : MD\SY
 Sample : J3562-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 W-14

Quant Time: Jun 18 04:58:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1697971	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2660915	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2249062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	727282	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	1099269	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
35) Dibromofluoromethane	7.59	113	986785	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
50) Toluene-d8	10.09	98	3718993	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
62) 4-Bromofluorobenzene	12.40	95	1019873	37.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.26%	
Target Compounds						
16) Acetone	3.82	43	10478	1.92	ug/l	99
18) Methyl Acetate	4.33	43	28393	1.70	ug/l	92
30) Chloroform	7.37	83	620932	15.88	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	16620	0.33	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049326.D
Acq On : 16 Jun 2018 22:29
Operator : MD\SY
Sample : J3562-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
W-14

Quant Time: Jun 18 04:58:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

