

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049327.D
 Acq On : 16 Jun 2018 22:55
 Operator : MD\SY
 Sample : J3562-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 W-12

Quant Time: Jun 18 07:18:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1628863	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2576515	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2161661	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	656693	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	1065754	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
35) Dibromofluoromethane	7.59	113	963253	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
50) Toluene-d8	10.09	98	3570501	44.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.96%	
62) 4-Bromofluorobenzene	12.40	95	956107	36.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.85	85	12723	0.68	ug/l	95
18) Methyl Acetate	4.33	43	24280	1.52	ug/l #	70
30) Chloroform	7.37	83	165053	4.40	ug/l	97
32) 1,1,1-Trichloroethane	7.57	97	24151	0.75	ug/l #	53
44) Trichloroethene	8.83	130	32163	1.35	ug/l	88
64) Tetrachloroethene	10.63	164	7141	0.32	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049327.D
Acq On : 16 Jun 2018 22:55
Operator : MD\SY
Sample : J3562-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
W-12

Quant Time: Jun 18 07:18:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

