

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
 Data File : VN049329.D
 Acq On : 16 Jun 2018 23:47
 Operator : MD\SY
 Sample : J3435-09 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NW6

Manual Integrations
 APPROVED

MMDadoda
 6/18/2018 3:16:31 PM

Quant Time: Jun 18 05:08:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1409351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2054812	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1740500	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	742956	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	837991	44.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.66%	
35) Dibromofluoromethane	7.59	113	754516	45.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.78%	
50) Toluene-d8	10.09	98	2882190	45.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.06%	
62) 4-Bromofluorobenzene	12.40	95	893918	42.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.42%	

Target Compounds

						Qvalue
18) Methyl Acetate	4.34	43	24788	1.79	ug/l	# 86
31) Cyclohexane	7.65	56	488590	17.56	ug/l	# 89
39) Methylcyclohexane	9.08	83	191541	7.58	ug/l	95
40) Benzene	8.04	78	431335	5.97	ug/l	98
52) Toluene	10.16	92	358289	8.46	ug/l	96
67) Ethyl Benzene	11.51	91	6519212	87.66	ug/l	98
68) m/p-Xylenes	11.62	106	6599221	236.44	ug/l	97
69) o-Xylene	11.95	106	1829316	68.05	ug/l	97
73) Isopropylbenzene	12.25	105	273133	4.19	ug/l	98
78) n-propylbenzene	12.59	91	610642	8.53	ug/l	100
80) 1,3,5-Trimethylbenzene	12.73	105	2246938	43.82	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	5830047	114.46	ug/l	98
85) sec-Butylbenzene	13.17	105	33901m	0.56	ug/l	
95) Naphthalene	15.14	128	686112	32.60	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061618\
Data File : VN049329.D
Acq On : 16 Jun 2018 23:47
Operator : MD\SY
Sample : J3435-09 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NW6

Manual Integrations
APPROVED

MMDadoda
6/18/2018 3:16:31 PM

Quant Time: Jun 18 05:08:51 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

