





Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018


1,2,3-Trichlorobenzene

Response Ratio




Resp Ratio = $9.01 \times 10^{-1} \times \text{Amt} - 6.23 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018




Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018



Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018


1,2,4-Trichlorobenzene

Response Ratio



Resp Ratio = 9.47×10^{-1} * Amt - 8.99×10^{-2}
Coef of Det (r^2) = 0.995 Curve Fit: Linear


Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018



Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018

1,1,2,2-Tetrachloroethane

Response Ratio




Resp Ratio = 8.63e-001 * Amt + 4.05e-002
Coef of Det (r²) = 1.000 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018

2-Chloroethyl Vinyl ether

Response Ratio



Resp Ratio = $1.68 \times 10^{-1} \times \text{Amt} - 4.82 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Calibration Table Last Updated: Mon Jun 18 15:25:38 2018