

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
 Data File : VN072922.D
 Acq On : 18 Jun 2022 07:26
 Operator : JC\MD
 Sample : VN0618WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0618WBL01

Quant Time: Jun 20 08:00:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	336128	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	567036	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	478476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	200179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	220983	48.628	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.260%
35) Dibromofluoromethane	7.945	113	171124	48.826	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.660%
50) Toluene-d8	10.375	98	646869	49.601	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.663	95	216858	46.243	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
Data File : VN072922.D
Acq On : 18 Jun 2022 07:26
Operator : JC\MD
Sample : VN0618WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0618WBL01

Quant Time: Jun 20 08:00:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

