

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
 Data File : VN072928.D
 Acq On : 18 Jun 2022 10:29
 Operator : JC\MD
 Sample : N3334-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-5-GR

Quant Time: Jun 20 08:02:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	304182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	536253	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	469385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	208828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	164287	39.949	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	79.900%		
35) Dibromofluoromethane	7.951	113	127865	38.577	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	77.160%		
50) Toluene-d8	10.374	98	442619	35.888	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	71.780%#		
62) 4-Bromofluorobenzene	12.668	95	155536	35.071	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	70.140%#		
Target Compounds					Qvalue	
16) Acetone	4.228	43	50013	22.124	ug/l	96
17) Carbon Disulfide	4.463	76	13864	1.858	ug/l #	91
20) Methylene Chloride	5.022	84	34608	9.675	ug/l	97
25) 2-Butanone	7.275	43	14861	4.143	ug/l	98
43) Isopropyl Acetate	8.492	43	42053	4.101	ug/l #	88
84) 1,2,4-Trimethylbenzene	13.298	105	38279	2.896	ug/l	99
95) Naphthalene	15.433	128	61481	3.811	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
Data File : VN072928.D
Acq On : 18 Jun 2022 10:29
Operator : JC\MD
Sample : N3334-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-5-GR

Quant Time: Jun 20 08:02:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

