

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
 Data File : VN072939.D
 Acq On : 18 Jun 2022 14:56
 Operator : JC\MD
 Sample : N3304-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-10-G

Quant Time: Jun 20 08:04:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	332276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	537161	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	488662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	242091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	241400	53.737	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.480%	
35) Dibromofluoromethane	7.957	113	202885	61.107	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 122.220%	
50) Toluene-d8	10.375	98	596392	48.274	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.540%	
62) 4-Bromofluorobenzene	12.669	95	256680	57.779	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.560%	
Target Compounds						
						Qvalue
16) Acetone	4.228	43	94983	38.464	ug/l	92
20) Methylene Chloride	5.022	84	36090	9.225	ug/l	93
25) 2-Butanone	7.269	43	35085	8.954	ug/l	92
43) Isopropyl Acetate	8.498	43	37388	3.640	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
Data File : VN072939.D
Acq On : 18 Jun 2022 14:56
Operator : JC\MD
Sample : N3304-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-10-G

Quant Time: Jun 20 08:04:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

