

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
 Data File : VN072940.D
 Acq On : 18 Jun 2022 15:20
 Operator : JC\MD
 Sample : N3304-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-9

Quant Time: Jun 20 08:05:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	326386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	528082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	475517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	223819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	243001	55.069	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.140%	
35) Dibromofluoromethane	7.951	113	202112	61.921	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 123.840%	
50) Toluene-d8	10.375	98	587733	48.391	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.780%	
62) 4-Bromofluorobenzene	12.669	95	246477	56.436	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.880%	
Target Compounds						
						Qvalue
16) Acetone	4.234	43	91094	37.555	ug/l	99
20) Methylene Chloride	5.022	84	57280	15.056	ug/l	94
25) 2-Butanone	7.275	43	36746	9.547	ug/l	96
43) Isopropyl Acetate	8.498	43	36635	3.628	ug/l #	88
84) 1,2,4-Trimethylbenzene	13.304	105	19578	1.382	ug/l	90
95) Naphthalene	15.427	128	39545	2.287	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
Data File : VN072940.D
Acq On : 18 Jun 2022 15:20
Operator : JC\MD
Sample : N3304-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-9

Quant Time: Jun 20 08:05:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

