

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
 Data File : VN072941.D
 Acq On : 18 Jun 2022 15:45
 Operator : JC\MD
 Sample : N3304-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-6-G

Quant Time: Jun 20 08:05:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	321564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	519218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	467623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	233191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	239098	54.997	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.000%
35) Dibromofluoromethane	7.951	113	198375	61.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	123.620%
50) Toluene-d8	10.374	98	580673	48.626	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.260%
62) 4-Bromofluorobenzene	12.668	95	249155	58.023	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.040%
Target Compounds						
						Qvalue
16) Acetone	4.234	43	87533	36.628	ug/l	88
18) Methyl Acetate	4.792	43	11121	1.748	ug/l	90
20) Methylene Chloride	5.010	84	28772	7.557	ug/l	93
25) 2-Butanone	7.269	43	13937	3.675	ug/l #	86
43) Isopropyl Acetate	8.492	43	35892	3.615	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061822\
Data File : VN072941.D
Acq On : 18 Jun 2022 15:45
Operator : JC\MD
Sample : N3304-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-6-G

Quant Time: Jun 20 08:05:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

