

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049388.D
 Acq On : 19 Jun 2018 16:29
 Operator : MD\SY
 Sample : PB110363ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110363ZHE#01

Quant Time: Jun 20 04:54:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1261913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1948387	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1638350	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	442486	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	841747	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
35) Dibromofluoromethane	7.59	113	733765	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
50) Toluene-d8	10.09	98	2703379	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
62) 4-Bromofluorobenzene	12.41	95	664403	33.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.96%	
Target Compounds						
16) Acetone	3.82	43	48675	11.978	ug/l #	89
18) Methyl Acetate	4.34	43	91126	7.354	ug/l	93
20) Methylene Chloride	4.55	84	32706	1.927	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049388.D
Acq On : 19 Jun 2018 16:29
Operator : MD\SY
Sample : PB110363ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110363ZHE#01

Quant Time: Jun 20 04:54:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

