

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049398.D
 Acq On : 19 Jun 2018 20:47
 Operator : MD\SY
 Sample : PB110363ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110363ZHE#10

Quant Time: Jun 20 05:22:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1233420	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1997067	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1634033	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	461488	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	840934	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
35) Dibromofluoromethane	7.59	113	735523	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
50) Toluene-d8	10.09	98	2694116	43.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.58%	
62) 4-Bromofluorobenzene	12.41	95	674091	33.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.28%	
Target Compounds						
16) Acetone	3.82	43	45035	11.338	ug/l	95
18) Methyl Acetate	4.33	43	119962	9.905	ug/l	97
20) Methylene Chloride	4.55	84	31933	1.925	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049398.D
Acq On : 19 Jun 2018 20:47
Operator : MD\SY
Sample : PB110363ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110363ZHE#10

Quant Time: Jun 20 05:22:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

