

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049399.D
 Acq On : 19 Jun 2018 21:13
 Operator : MD\SY
 Sample : PB110363ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110363ZHE#11

Quant Time: Jun 20 05:24:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1195644	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1912259	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1560891	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	439282	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	817767	51.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.14%	
35) Dibromofluoromethane	7.59	113	715689	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
50) Toluene-d8	10.09	98	2582889	43.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.70%	
62) 4-Bromofluorobenzene	12.40	95	643775	33.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.10%	
Target Compounds						
16) Acetone	3.82	43	44487	11.554	ug/l #	89
18) Methyl Acetate	4.33	43	120081	10.228	ug/l	97
20) Methylene Chloride	4.55	84	34868	2.168	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049399.D
Acq On : 19 Jun 2018 21:13
Operator : MD\SY
Sample : PB110363ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110363ZHE#11

Quant Time: Jun 20 05:24:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

