

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049404.D
 Acq On : 19 Jun 2018 23:21
 Operator : MD\SY
 Sample : PB110363ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110363ZHE#16

Quant Time: Jun 20 05:42:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1195882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1939401	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1560942	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	414159	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	812420	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
35) Dibromofluoromethane	7.59	113	714269	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
50) Toluene-d8	10.09	98	2594866	43.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.86%	
62) 4-Bromofluorobenzene	12.41	95	628927	31.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.68%	
Target Compounds						
16) Acetone	3.83	43	43488	11.292	ug/l	93
18) Methyl Acetate	4.33	43	107568	9.160	ug/l	93
20) Methylene Chloride	4.54	84	32092	1.995	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049404.D
Acq On : 19 Jun 2018 23:21
Operator : MD\SY
Sample : PB110363ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110363ZHE#16

Quant Time: Jun 20 05:42:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

