

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049413.D
 Acq On : 20 Jun 2018 4:04
 Operator : MD\SY
 Sample : PB110363ZHE#20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110363ZHE#20

Quant Time: Jun 20 08:38:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1186278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1830822	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1511049	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	408376	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	774008	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
35) Dibromofluoromethane	7.59	113	692146	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
50) Toluene-d8	10.09	98	2524900	44.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.54%	
62) 4-Bromofluorobenzene	12.41	95	624911	33.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.02%	
Target Compounds						
16) Acetone	3.83	43	42271	11.065	ug/l	93
18) Methyl Acetate	4.33	43	81000	6.954	ug/l	96
20) Methylene Chloride	4.55	84	32347	2.027	ug/l	86
95) Naphthalene	15.14	128	16674	3.548	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049413.D
Acq On : 20 Jun 2018 4:04
Operator : MD\SY
Sample : PB110363ZHE#20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110363ZHE#20

Quant Time: Jun 20 08:38:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

