

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049416.D
 Acq On : 20 Jun 2018 5:21
 Operator : MD\SY
 Sample : J3575-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11-S

Quant Time: Jun 20 07:40:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1283377	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2048913	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1693476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	501468	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	856034	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
35) Dibromofluoromethane	7.59	113	761755	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
50) Toluene-d8	10.09	98	2784860	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
62) 4-Bromofluorobenzene	12.40	95	716134	34.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.62%	
Target Compounds						
16) Acetone	3.82	43	46634	11.283	ug/l	93
18) Methyl Acetate	4.33	43	89533	7.105	ug/l	93
20) Methylene Chloride	4.55	84	11282618	653.631	ug/l	91
95) Naphthalene	15.14	128	194599	15.546	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049416.D
Acq On : 20 Jun 2018 5:21
Operator : MD\SY
Sample : J3575-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11-S

Quant Time: Jun 20 07:40:24 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

