

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049417.D
 Acq On : 20 Jun 2018 5:46
 Operator : MD\SY
 Sample : J3573-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-53-COMP

Quant Time: Jun 20 08:26:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1176624	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1891000	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1541698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	459483	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	794022	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
35) Dibromofluoromethane	7.59	113	699649	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
50) Toluene-d8	10.09	98	2559419	43.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.88%	
62) 4-Bromofluorobenzene	12.40	95	652405	33.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.74%	
Target Compounds						
16) Acetone	3.82	43	47173	12.449	ug/l	92
18) Methyl Acetate	4.33	43	89704	7.764	ug/l	92
20) Methylene Chloride	4.55	84	105435	6.662	ug/l	91
44) Trichloroethene	8.84	130	26155	1.491	ug/l	84
68) m/p-Xylenes	11.62	106	29019	1.174	ug/l	96
95) Naphthalene	15.14	128	798955	56.000	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049417.D
Acq On : 20 Jun 2018 5:46
Operator : MD\SY
Sample : J3573-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-53-COMP

Quant Time: Jun 20 08:26:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

