

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049418.D
 Acq On : 20 Jun 2018 6:12
 Operator : MD\SY
 Sample : J3573-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-52-COMP

Quant Time: Jun 20 08:29:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1240619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1964544	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1626354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	473848	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	809171	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
35) Dibromofluoromethane	7.59	113	731259	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
50) Toluene-d8	10.09	98	2693695	44.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.02%	
62) 4-Bromofluorobenzene	12.40	95	688755	34.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.84%	
Target Compounds						
16) Acetone	3.82	43	45423	11.369	ug/l	98
18) Methyl Acetate	4.33	43	92160	7.565	ug/l	97
20) Methylene Chloride	4.55	84	229476	13.752	ug/l	91
68) m/p-Xylenes	11.62	106	63689	2.442	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	62223	1.915	ug/l	98
95) Naphthalene	15.14	128	204369	16.983	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049418.D
Acq On : 20 Jun 2018 6:12
Operator : MD\SY
Sample : J3573-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-52-COMP

Quant Time: Jun 20 08:29:04 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

