

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049421.D
 Acq On : 20 Jun 2018 7:28
 Operator : MD\SY
 Sample : J3249-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-01-061818-D

Quant Time: Jun 20 08:37:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1127753	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1838619	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1519418	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	448872	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	763770	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
35) Dibromofluoromethane	7.59	113	675584	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
50) Toluene-d8	10.09	98	2489270	43.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.90%	
62) 4-Bromofluorobenzene	12.40	95	645628	34.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.94%	
Target Compounds						
16) Acetone	3.82	43	36357	10.011	ug/l	99
18) Methyl Acetate	4.33	43	106484	9.616	ug/l	95
20) Methylene Chloride	4.55	84	53875	3.552	ug/l #	83
95) Naphthalene	15.14	128	19269	3.623	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049421.D
Acq On : 20 Jun 2018 7:28
Operator : MD\SY
Sample : J3249-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-01-061818-D

Quant Time: Jun 20 08:37:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

