

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049427.D
 Acq On : 20 Jun 2018 10:02
 Operator : MD\SY
 Sample : J3573-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-49-COMP

Quant Time: Jun 20 11:45:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1141463	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1857607	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1503549	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	510835	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	764304	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
35) Dibromofluoromethane	7.59	113	689037	45.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.70%	
50) Toluene-d8	10.09	98	2491306	43.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.08%	
62) 4-Bromofluorobenzene	12.40	95	662797	35.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.06%	
Target Compounds						
16) Acetone	3.82	43	58915	16.027	ug/l	92
17) Carbon Disulfide	4.05	76	51907	1.265	ug/l	99
20) Methylene Chloride	4.55	84	2235248	145.593	ug/l	91
84) 1,2,4-Trimethylbenzene	13.04	105	36527	1.043	ug/l	100
95) Naphthalene	15.14	128	826516	52.693	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049427.D
Acq On : 20 Jun 2018 10:02
Operator : MD\SY
Sample : J3573-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-49-COMP

Quant Time: Jun 20 11:45:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

