

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061918\
 Data File : VN049433.D
 Acq On : 20 Jun 2018 12:36
 Operator : MD\SY
 Sample : J3500-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-55(0-5)

Quant Time: Jun 20 13:56:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1193189	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1921153	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1560497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	432444	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	803316	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
35) Dibromofluoromethane	7.59	113	708330	45.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.14%	
50) Toluene-d8	10.09	98	2573668	43.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.98%	
62) 4-Bromofluorobenzene	12.40	95	651780	33.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.62%	
Target Compounds						
16) Acetone	3.82	43	62233	16.196	ug/l	92
20) Methylene Chloride	4.55	84	95826	5.971	ug/l	89
51) 4-Methyl-2-Pentanone	9.99	43	16027	1.337	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061918\
Data File : VN049433.D
Acq On : 20 Jun 2018 12:36
Operator : MD\SY
Sample : J3500-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-55(0-5)

Quant Time: Jun 20 13:56:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

