

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062018\
 Data File : VN049442.D
 Acq On : 20 Jun 2018 17:55
 Operator : MD\SY
 Sample : J3575-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Quant Time: Jun 21 02:02:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1197707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1884182	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1550232	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	477799	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	799046	50.30	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.60%	
35) Dibromofluoromethane	7.59	113	697386	45.75	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.50%	
50) Toluene-d8	10.09	98	2561583	44.13	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.26%	
62) 4-Bromofluorobenzene	12.40	95	665454	34.67	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	69.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	56521	14.65	ug/l	100
18) Methyl Acetate	4.34	43	90964	7.73	ug/l	93
20) Methylene Chloride	4.55	84	83142	5.16	ug/l	87
95) Naphthalene	15.14	128	144543	12.67	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062018\
Data File : VN049442.D
Acq On : 20 Jun 2018 17:55
Operator : MD\SY
Sample : J3575-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Jun 21 02:02:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

