

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062018\
 Data File : VN049443.D
 Acq On : 20 Jun 2018 18:21
 Operator : MD\SY
 Sample : J3581-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SSP-B-COMPOSITE

Quant Time: Jun 21 02:12:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1142351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1802763	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1477125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	504886	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	774573	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
35) Dibromofluoromethane	7.59	113	679486	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
50) Toluene-d8	10.09	98	2432231	43.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.60%	
62) 4-Bromofluorobenzene	12.40	95	649679	35.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.76%	
Target Compounds						
16) Acetone	3.82	43	45537	12.38	ug/l	96
18) Methyl Acetate	4.33	43	90325	8.05	ug/l	95
20) Methylene Chloride	4.55	84	176459	11.48	ug/l	89
95) Naphthalene	15.14	128	62020	6.46	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062018\
Data File : VN049443.D
Acq On : 20 Jun 2018 18:21
Operator : MD\SY
Sample : J3581-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SSP-B-COMPOSITE

Quant Time: Jun 21 02:12:48 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

