

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062018\
 Data File : VN049447.D
 Acq On : 20 Jun 2018 20:04
 Operator : MD\SY
 Sample : J3245-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-061918-B

Quant Time: Jun 21 02:20:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1166761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1868419	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1550707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	517590	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	790028	51.05	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.10%	
35) Dibromofluoromethane	7.59	113	694414	45.94	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.88%	
50) Toluene-d8	10.09	98	2543481	44.19	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.38%	
62) 4-Bromofluorobenzene	12.41	95	691613	36.34	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	72.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.41	74	13173	1.64	ug/l	94
16) Acetone	3.83	43	35665	9.49	ug/l	97
18) Methyl Acetate	4.33	43	104092	9.09	ug/l	93
20) Methylene Chloride	4.55	84	89579	5.71	ug/l	90
30) Chloroform	7.38	83	27290	1.02	ug/l	87
44) Trichloroethene	8.84	130	17364	1.00	ug/l	88
95) Naphthalene	15.14	128	595244	39.37	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062018\
Data File : VN049447.D
Acq On : 20 Jun 2018 20:04
Operator : MD\SY
Sample : J3245-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-061918-B

Quant Time: Jun 21 02:20:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

