

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062018\
 Data File : VN049454.D
 Acq On : 20 Jun 2018 23:04
 Operator : MD\SY
 Sample : J3584-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1-D

Quant Time: Jun 21 03:15:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1207430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1931130	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1565723	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	482899	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	812335	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
35) Dibromofluoromethane	7.59	113	704251	45.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.16%	
50) Toluene-d8	10.09	98	2607068	43.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.64%	
62) 4-Bromofluorobenzene	12.41	95	676690	34.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	43207	11.11	ug/l	# 82
18) Methyl Acetate	4.33	43	106542	8.99	ug/l	97
20) Methylene Chloride	4.55	84	1699698	104.66	ug/l	92
64) Tetrachloroethene	10.63	164	55153	3.44	ug/l	98
95) Naphthalene	15.14	128	147354	12.76	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062018\
Data File : VN049454.D
Acq On : 20 Jun 2018 23:04
Operator : MD\SY
Sample : J3584-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1-D

Quant Time: Jun 21 03:15:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

